



**RADWAG WAGI ELEKTRONICZNE**  
**WITOLD LEWANDOWSKI**  
POLAND, 26-600 Radom, 28 Bracka Street  
tel. 048/384 88 00 · fax 048/385 00 10  
e-mail: radom@radwag.pl · http://www.radwag.pl

---

## CONFORMITY DECLARATION

**Manufacturer:** **RADWAG WAGI ELEKTRONICZNE**  
**Witold Lewandowski**

**Address:** POLAND, 26-600 Radom, 28 Bracka Street

**The medical device:**

Baby scales **WPT 6/15D ; WPT 10/20D**  
Personal scales **WPT 60/150 O ; WPT 60/150 OW ; WPT 100/200 O ; WPT 100/200 OW**  
Bed scales **WPT/4B 500C ; WPT/8B 300C**  
Chair scales **WPT/4K 150C**

Covered by the Technical Files rev.

Baby scales KW 0003, dated 27.03.2007  
Personal scales KW 0002, dated 22.05.2007  
Bed scales KW 0020, dated 24.09.2008 to WPT/4B 500C  
KW 0035, dated 24.09.2009 to WPT/8B 300C  
Chair scales KW 0027, dated 24.09.2008

**Class I with measuring function, rule 12 Annex IX of the 93/42/EEC directive**

**We hereby declare in sole responsibility, that the designated product, fulfils the safety requirements of the European Directives:**

Directive 90/384/EEC  
Directive 93/42/EEC

**Conformity assessment procedure according to MDD directive – Annex VII + Annex V**

**Additional information:**

- Type Approval Certificate EC no. T6496 issued by  
Notified Body no. 0122 NEDERLANDS MEETINSTITUUT
- Test Certificate for measuring indicator no. ZR 128/06-0050 issued by  
Notified Body no. 1383 ČESKÝ METROLOGICKÝ INSTITUT
- Conformity with Directive 90/384/EEC- ČESKÝ METROLOGICKÝ INSTITUT no 1383
- Conformity with Directive 93/42/EEC- TÜV Rheinland Product Safety GmbH no 0197  
Certification Office  
Am Grauen Stein  
51105 Cologne

With the authorisation of:

The Managing Director of RADWAG WAGI ELEKTRONICZNE – Dr. Eng Witold  
Lewandowski

*Radom 01.10.2009*

.....  
*Place and date of issue*

*Attorney for QMS Jacek Pilecki*

.....  
*Position, Name*

*J. Pilecki*  
.....  
*Signature*